



HPLC Analysis for V1301401 Deslorelin

Data file: D:\CHEM32\1\DATA\250321\800008_25020180Final.D
Sample name: 800008_25020180Final
Description: Mobile Phase: Buffer A: 0.1% TFA in H₂O
Buffer B: 0.1% TFA in ACN
Gradient: 27-37%B in 20min, Flow rate: 1.0ml/min
Column: Unitary C18 5um 100A 4.6*150mm
Instrument: Agilent HPLC 1100
Instrument: 1100 Location: 66
Injection date: 2025/3/21 13:19:31 Injection: 1 of 1
Last changed: 2025/3/21 12:35:32 Injection volume: 5.000
Acq. operator: SYSTEM



Signal: DAD1 A, Sig=220, 4 Ref=360, 100

RT [min]	Type	Width [min]	Area	Height	Area %	Name
5.899	BB	0.1260	4.9535	0.6423	0.0273	
7.246	MM	0.3052	58.2701	3.1825	0.3208	
8.036	MF	0.2020	5.5887	0.4611	0.0308	
8.443	FM	0.2227	18077.2324	1353.1851	99.5276	
9.033	FM	0.1459	16.9876	1.9402	0.0935	
Sum			18163.0323			



Product Name: Deslorelin
InvivoChem Cat #: V13014
CAS #: 57773-65-6
Batch #: V1301401
Formula: C₆₄H₈₃N₁₇O₁₂
Molecular Weight: 1282.4505