

HPLC Analysis for V25774 Yatein

Analysis Report

<Sample Information>


Sample Name : CFN98635
Instrument : SHIMADZU LC-20AT
Column : InertSustain-C18(5 µm, 4.6×250 mm)
Column Temp. : 35°C
Mobile phase flow : 1.0 ml/min
Injection Volume : 5 µL
Detection Wave. : UV-254nm
Con. & Solvent : 0.5mg/ml (Methanol)

<Analysis condition>

5%→100% Acetonitrile in 0.05% Phosphoric acid H2O
over 10.0min followed by 100% Acetonitrile to 17.0 min

<Chromatogram>

mV



<Peak Table>

Detector A 254nm

Peak#	Ret. Time	Area	Height	Conc.	Unit	Area%
1	14.616	508594	137178	1.967		1.967
2	15.523	25341347	3352284	98.033		98.033
Total		25849941	3489462			100.000