



HPLC Analysis for V3800001 RO27-3225 TFA

Data file: C:\CHEM32\1\DATA\210316\20210316 2021-03-17 11-13-19\603246 21030014Final.D
Sample name: 603246 21030014Final
Description: Mobile phase: Buffer A: 0.1%TFA in H2O
Buffer B: 0.1%TFA in ACN
Gradient: 22-32% B in 20min, Flow rate: 1.0ml/min
Column: phenomenex C18, 4.6*150mm, 5um 100A (A-0013)
Instrument: Agilent HPLC 1100
Instrument: 1100 Location: 72
Injection date: 2021/3/17 11:43:20 Injection: 1 of 1
Last changed: 2021/3/17 12:04:07 Injection volume: 10.000
Acq. operator: qc



Signal: DAD1 A, Sig=220, 4 Ref=360, 100

RT[mi n]	Type	Width[mi n]	Area	Height	Area%	Name
7.651	BB	0.2758	37.5114	1.9557	0.0836	
8.623	BB	0.3898	319.2060	10.8943	0.7113	
9.694	BV E	0.1719	29.4025	2.7696	0.0655	
10.403	VB R	0.4294	44490.1133	1677.4441	99.1396	
	Sum		44876.2332			



Product Name: R027-3225 TFA
InvivoChem Cat #: V38000
CAS #: 1373926-49-8
Batch #: V3800001
Formula: C₄₁H₅₃F₃N₁₂O₈
Molecular Weight: 898.93